



More information on the website  
[mirror.radwag.com/en/info,w1,9XE](https://mirror.radwag.com/en/info,w1,9XE)

# 100 $\mu$ l ÷ 1000 $\mu$ l - RW8-106-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

Metrological parameters	
Volume	100 – 1000 $\mu$ l
Systematic error	2.0% / 0.6%
Random error	0.7% / 0.2%
Tips	1000 $\mu$ l



## Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration