



More information on the website  
[mirror.radwag.com/en/info,w1,A7T](http://mirror.radwag.com/en/info,w1,A7T)

# 2 $\mu$ l ÷ 20 $\mu$ l - RW8-102-20-9 Automatic Pipettes

PA-203-0002



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

### Metrological parameters

|                  |                |
|------------------|----------------|
| Volume           | 2 – 20 $\mu$ l |
| Systematic error | 3.0% / 0.9%    |
| Random error     | 2.0% / 0.4%    |
| Tips             | 300 $\mu$ l    |



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for Pipettes Calibration  
 Pipette Tips