



More information on the website
mirror.radwag.com/it/info,w1,A7T

Pipetta automatica 2 µl ÷ 20 µl - RW8-102-20-9

PA-203-0002



The drawings, photos and graphics used are for illustrative purposes only.

Specifiche

Parametri metrologici

Volume	2 – 20 µl
Systematic error	3,0% / 0,9%
Random error	2,0% / 0,4%
Tips	300 µl



Compatibile con (Additional Fee)

Microbilancia per la calibrazione delle pipette XA 5Y.M.A.P
 Microbilancia per la calibrazione delle pipette MYA 5Y.P

Bilance automatiche per la calibrazione delle pipette
 Puntali per pipette