



More information on the website
mirror.radwag.com/us/info,w1,A7T

2 μ l ÷ 20 μ l - RW8-102-20-9 Automatic Pipettes

PA-203-0002



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters

Volume	2 – 20 μ l
Systematic error	3,0% / 0,9%
Random error	2,0% / 0,4%
Tips	300 μ l



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration
 Pipette Tips