



More information on the website
mirror.radwag.com/en/info,w1,C80

10 μ l ÷ 100 μ l - RW8-104-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	10 – 100 μ l
Systematic error	3.0% / 0.8%
Random error	1.5% / 0.15%
Tips	300 μ l



Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration
Pipette Tips