10 µl ÷ 100 µl - RW8-104-20-9 Automatic Pipettes

PA-203-0004

More information on the website
mirror.radwag.com/en/info,w1,C80

The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	10 – 100 µl
Systematic error	3.0% / 0.8%
Random error	1.5% / 0.15%
Tips	300 µl

2 YEARS WARRANTY

Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration
Automatic Balances for Pipettes Calibration
Pipette Tips