



More information on the website  
[mirror.radwag.com/en/info,w1,IRH](http://mirror.radwag.com/en/info,w1,IRH)

# 20 $\mu$ l ÷ 200 $\mu$ l - RW8-105-20-9 Automatic Pipettes

PA-203-0005



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

### Metrological parameters

Volume	20 – 200 $\mu$ l
Systematic error	2.0% / 0.6%
Random error	0.8% / 0.15%
Tips	300 $\mu$ l



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for Pipettes Calibration