



More information on the website
mirror.radwag.com/it/info,w1,IRH

Pipetta automatica 20 μ l ÷ 200 μ l - RW8-105-20-9

PA-203-0005



The drawings, photos and graphics used are for illustrative purposes only.

Specifiche

Parametri metrologici	
Volume	20 – 200 μ l
Systematic error	2,0% / 0,6%
Random error	0,8% / 0,15%
Tips	300 μ l



Compatibile con (Additional Fee)

Microbilancia per la calibrazione delle pipette XA 5Y.M.A.P
 Microbilancia per la calibrazione delle pipette MYA 5Y.P

Bilance automatiche per la calibrazione delle pipette