



More information on the website  
[mirror.radwag.com/us/info,w1,IRH](https://mirror.radwag.com/us/info,w1,IRH)

# 20 $\mu$ l ÷ 200 $\mu$ l - RW8-105-20-9 Automatic Pipettes

PA-203-0005



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

| Metrological parameters |                  |
|-------------------------|------------------|
| Volume                  | 20 – 200 $\mu$ l |
| Systematic error        | 2,0% / 0,6%      |
| Random error            | 0,8% / 0,15%     |
| Tips                    | 300 $\mu$ l      |



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration