radwag.com

5 µl ÷ 50 µl - RW8-103-20-9 Automatic Pipettes

PA-203-0003

More information on the website
mirror.radwag.com/en/info,w1,QLH

The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	5 – 50 µl
Systematic error	2.0% / 0.6%
Random error	2.0% / 0.3%
Tips	300 µl

Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration
Automatic Balances for Pipettes Calibration
Pipette Tips