



More information on the website  
[mirror.radwag.com/us/info,w1,QLH](https://mirror.radwag.com/us/info,w1,QLH)

# 5 µl ÷ 50 µl - RW8-103-20-9 Automatic Pipettes

PA-203-0003



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

Metrological parameters	
Volume	5 – 50 µl
Systematic error	2,0% / 0,6%
Random error	2,0% / 0,3%
Tips	300 µl



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration  
 Pipette Tips