



More information on the website
mirror.radwag.com/en/info,w1,UAW

0.5 µl ÷ 10 µl - RW8-101-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	0.5 – 10 µl
Systematic error	2.5% / 1.0%
Random error	1.5% / 0.8%
Tips	10 µl



Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration
 MYA 5Y.P Microbalance for pipettes calibration

Pipette Tips
 Automatic Balances for. Pipettes Calibration