



More information on the website
mirror.radwag.com/en/info,w1,VML

500 μ l ÷ 5000 μ l - RW8-107-20-9 Automatic Pipettes

PA-203-0007



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters

Volume	500 – 5000 μ l
Systematic error	2.0% / 0.5%
Random error	0.6% / 0.15%
Tips	5000 μ l



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for Pipettes Calibration