radwag.com

500 µl ÷ 5000 µl - RW8-107-20-9 Automatic Pipettes

PA-203-0007

More information on the website
mirror.radwag.com/en/info,w1,VML



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	500 – 5000 µl
Systematic error	2.0% / 0.5%
Random error	0.6% / 0.15%
Tips	5000 µl

Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for Pipettes Calibration