



More information on the website
mirror.radwag.com/it/info,w1,VML

Pipetta automatica 500 μ l ÷ 5000 μ l - RW8-107-20-9

PA-203-0007



The drawings, photos and graphics used are for illustrative purposes only.

Specifiche

Parametri metrologici	
Volume	500 – 5000 μ l
Systematic error	2,0% / 0,5%
Random error	0,6% / 0,15%
Tips	5000 μ l



Compatibile con (Additional Fee)

Microbilancia per la calibrazione delle pipette XA 5Y.M.A.P
 Microbilancia per la calibrazione delle pipette MYA 5Y.P

Bilance automatiche per la calibrazione delle pipette